

Certificate

PROFIBUS Nutzerorganisation e.V. grants to

Endress + Hauser
Kaegenstrasse 7, CH-4153 Reinach, Switzerland

the Certificate No: **Z01726** for the PROFIBUS Device:

Model Name: Promass 200
Revision: Profile 3.02; SW/FW: 01.01; HW: 01.00
GSD: EH3x155F.GSD, File Version: 13.01.2015
pa039742.GSD

This certificate confirms that the product has successfully passed the certification tests with the following scope:

<input checked="" type="checkbox"/>	DP-V0	MS0, Sync, Freeze, Set_Slave_Add
<input checked="" type="checkbox"/>	DP-V1	MS1, Prm_Block_Structure, MS2, I&M
<input checked="" type="checkbox"/>	Profile	PROFIBUS PA 3.02
<input checked="" type="checkbox"/>	Physical Layer	MBP

Test Report Number: FZI0023
Authorized Test Laboratory: FZI Forschungszentrum Informatik, Karlsruhe, Germany

The tests were executed in accordance with the following documents:
 "Test Specifications for PROFIBUS DP Slaves, Version 3.0.9, September 2008" and
 "Test Specifications for PROFIBUS PA Profile 3.0.2, Version 5.0.4, February 2010".


This certificate is granted according to the document:
 "Framework for testing and certification of PROFIBUS and PROFINET products".

For all products that are placed in circulation by April 14, 2018 the certificate is valid for life.


Board of PROFIBUS Nutzerorganisation e. V.



(Official in Charge)



(Karsten Schneider)

(K.-P. Lindner)